

.

From: Clemmer, Jill (DPH)
Sent: Tuesday, November 01, 2011 3:16 PM
To: Servizio, Paul (DPH); Piro, Peter (DPH)
Cc: Nassif, Julianne (DPH)
Subject: Fentanyl fix

After wracking my head and a few others' heads today, I am convinced that our fentanyl fix lies in the "fix" of the calibration standards. One or both of the following must be done:

- 1) pH the std as the QC is, at 4.5
- 2) dissolve the standard in a small amount of methanol prior to water pH'd at 4.5

Let's make new and improved standards tomorrow... ok?
Jill